

Heartsine Technologies Ltd
 203 Airport Road West,
 Belfast, BT3 9ED,
 United Kingdom

Page 1 of 1

Certificate of Test

Electrical Safety for Automated External Defibrillator

Date of Issue: 11th February 2008

SGS Reference: ELS 150121/1/C/AC/08

Details of Sample: Model: PDU
 Rated: Type BF, Internal, non rechargeable, non replaceable battery

Date of Receipt: 20th December 2007

Date of Test: 20th December 2007 to 8th February 2008

Test Specification: EN 60601-1 : 1990 plus amendments A1 : 1993, All : 1993, A12 : 1993, A2 : 1995 and A13 : 1996,
 IEC 60601-1: 1988 including amendments 1 : 1991 and 2 : 1995
 and corrigendum 1995
 Safety of medical electrical equipment .

Conclusion: The sample tested was found to comply with the Test Specification.

This document is issued by the Company under its General Conditions of Service printed overleaf or available on request and accessible at http://www.sgs.com/terms_and_conditions.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 28 days only. This document cannot be reproduced except in full, without prior approval of the Company.

Tests marked * are not covered under our UKAS scope

Test Engineer/ Authorised Signatory

A Clifford



Authorised Signatory/ Counter Signatory:

R Gough



All enquires relating to this certificate must be directed to the Authorised Signatory