

To whom it may concern:

Dear Sirs

This is to confirm that the HeartSine Technologies samaritan PAD and AED do not have any user serviceable parts and therefore the only maintenance required is to ensure that the status indicators are not showing any warnings of failures of the self test, and that the battery and defibrillation electrodes are within the relevant expiry dates.

In addition due to the design and manufacture there is not requirement to carry out an annual test of the defibrillation energy output.

Yours faithfully



Dr Allister McIntyre  
Senior Design Engineer