

Heartsine Technologies Ltd  
 Canberra House  
 203 Airport Road West  
 Belfast  
 Ireland  
 BT3 9ED

Page 1 of 1

## Certificate of Test

### Electrical Safety

#### For

## Portable Defibrillator Unit

**Date of Issue:** 18<sup>th</sup> August 2006

**SGS Reference:** ELS 102384/1/C/RG/06

**Details of Sample:** Model : Samaritan Pad "SAMPAD" P/N - SAM 300  
 Serial No : 04-0001356

**Date of Receipt:** 17<sup>th</sup> July 2006

**Date of Test:** 20<sup>th</sup> July 2006 to 11<sup>th</sup> August 2006

**Test Specification:** Test for compliance with IP56 as described in BS EN 60529:1992 incorporating Amendments 1 & 2 (Degrees of protection provided by enclosures (IP code))

**Conclusion:** The sample tested was found to comply with IP5X (category 1) and IPX6. and therefore complies with IP56 as required by the standard.


*This certificate is issued by the Company subject to its General Conditions of Service available upon request and accessible at [www.sgs.com](http://www.sgs.com). Attention is drawn to the limitations of liability, indemnification, jurisdictional, representative restrictions of test samples and sample retention policies defined therein.*

*This certificate refers only to the sample submitted for test.*

*This certificate shall not be reproduced except in full without the written approval of SGS United Kingdom Ltd.*

**Authorised Signatory:** R. N. Gough

**Counter Signatory :** S. Wilkinson




All enquires relating to this certificate must be directed to the Authorised Signatory

QUAL 08 Rev 5